

Sample Name :GnRH-I
 Cat No.: RP20303
 Lot No.: E9540720F
 Time Processed :8:42:38
 Month-Day-Year Processed :06/26/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>


Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
33.00	Pumps	Solvent B Conc.	5
33.00	Controller	Stop	

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment:ZJ19010324

<Chromatogram>



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	10.854	97079	12078	0.437
2	11.183	54788	7190	0.246
3	11.392	27252	3629	0.123
4	11.552	42702	5100	0.192
5	11.979	21824952	2378308	98.144
6	12.348	170611	15544	0.767
7	12.817	20309	1853	0.091
Total		22237693	2423702	100.000