

Sample Name :GRGDSPC  
 Cat No. :RP20283  
 Lot No. :E0847920H  
 Time Processed :6:22:45 PM  
 Month-Day-Year Processed :08/20/20

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
35.00	Pumps	Solvent B Conc.	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment: ZJ17010508

### <Chromatogram>

mV



### <Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	5.033	4845	1070	0.038
2	5.540	12543365	1085244	97.712
3	6.062	37154	6388	0.289
4	6.598	118299	19685	0.922
5	7.128	31864	5318	0.248
6	8.684	1414	242	0.011
7	11.597	2966	169	0.023
8	12.452	26319	912	0.205
9	15.950	1213	112	0.009
10	16.892	1955	170	0.015
11	17.275	4525	367	0.035
12	17.884	2296	161	0.018
13	18.282	1126	121	0.009
14	18.903	2910	290	0.023
15	19.392	4350	377	0.034

Peak#	Ret. Time	Area	Height	Area%
16	20.780	32166	2259	0.251
17	22.550	3907	286	0.030
18	23.364	10147	600	0.079
19	24.435	5159	763	0.040
20	24.873	1137	154	0.009
Total		12837117	1124688	100.000