

Sample Name : JAG-1, scrambled
 Cat No.: RP20525
 Lot No.: E5618820E
 Time Processed : 8:37:18
 Month-Day-Year Processed : 05/08/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

<<LC Time Program>>


| Time | Module | Command | Value |
|-------|------------|-----------------|-------|
| 0.01 | Pumps | Solvent B Conc. | 5 |
| 25.00 | Pumps | Solvent B Conc. | 65 |
| 25.01 | Pumps | Solvent B Conc. | 95 |
| 27.00 | Pumps | Solvent B Conc. | 95 |
| 27.01 | Pumps | Solvent B Conc. | 5 |
| 33.00 | Pumps | Solvent B Conc. | 5 |
| 33.00 | Controller | Stop | |

<<Column Performance>>

<Detector A>
 Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010324

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

| Peak# | Ret. Time | Area | Height | Area% |
|-------|-----------|---------|--------|---------|
| 1 | 5.161 | 10398 | 1843 | 0.293 |
| 2 | 13.262 | 135918 | 18143 | 3.831 |
| 3 | 13.422 | 3392631 | 566044 | 95.617 |
| 4 | 20.100 | 9187 | 1421 | 0.259 |
| Total | | 3548134 | 587450 | 100.000 |