

Sample Name :Pep-1 (uncapped)
 Cat. No. : RP10130
 Lot. No. : E3923119F
 Time Processed :7:36:33
 Month-Day-Year Processed :06/20/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
33.00	Pumps	Solvent B Conc.	5
33.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ17010508

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	16.648	19038	3498	0.120
2	17.145	15875	2708	0.100
3	17.911	88122	5138	0.554
4	18.263	55970	9520	0.352
5	18.476	15588871	2129655	98.082
6	18.818	95005	12091	0.598
7	19.616	30861	4764	0.194
Total		15893741	2167374	100.000