


Sample Name :HEL (46-61)
 Cat. No. : RP11736
 Lot. No. : PE1871807

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	5	
40.00	Pumps	Pump A B.Conc	5	
40.01	Controller	Stop		

<<Column Performance>>
 <Detector A>
 Column : Inertsil ODS-3 4.6 x 250 mm



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	5.671	14268	1969	0.765
2	12.667	3584	435	0.192
3	13.202	4313	425	0.231
4	13.544	60268	7247	3.230
5	13.773	1773071	280690	95.027
6	14.000	7140	859	0.383
7	14.340	3220	359	0.173
Total		1865866	291983	100.000