

Sample Name : N-Boc-Phe-Leu-Phe-Leu-Phe  
 Cat. No. : RP12950  
 Lot. No. : PE6081712

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow: 1 ml/min  
 Wavelength: 220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
10.00	Pumps	Pump A B.Conc	35	
35.00	Pumps	Pump A B.Conc	95	
41.00	Pumps	Pump A B.Conc	95	
41.01	Pumps	Pump A B.Conc	5	
50.00	Pumps	Pump A B.Conc	5	
50.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Chromatogram



1 Det. A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	11.074	18615	3871	0.465
2	20.450	13577	1678	0.339
3	20.968	3830986	531806	95.757
4	21.753	12475	2371	0.312
5	22.916	54726	8593	1.368
6	23.098	70354	9071	1.759
Total		4000734	557390	100.000