

Sample Name :MOG(35-55)
 Cat.No :RP10245
 Lot.No : E5759920C
 Time Processed: 4:46:37
 Year-Month-Day: 2020/3/27

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min

<<Detector A>>
 Wavelength Ch1 : 220 nm

<<LC Time Program>>


Time	Module	Command	Value	Comment
0.01	Pumps	Solvent B Conc.	5	
25.00	Pumps	Solvent B Conc.	65	
25.01	Pumps	Solvent B Conc.	95	
27.00	Pumps	Solvent B Conc.	95	
27.01	Pumps	Solvent B Conc.	5	
33.00	Pumps	Solvent B Conc.	5	
33.01	Controller	Stop		

<<Column Performance>>

<Detector A>
 Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment:ZJ19010015

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	12.359	4036	342	0.036
2	13.617	1037	218	0.009
3	13.950	18385	1536	0.165
4	14.118	41695	5051	0.375
5	14.348	69471	7463	0.625
6	14.573	246245	24745	2.215
7	14.811	46328	7181	0.417
8	15.044	10679534	1535585	96.079
9	15.622	1411	366	0.013
10	15.764	1521	420	0.014
11	15.929	1662	264	0.015
12	16.428	4082	288	0.037
Total		11115407	1583458	100.000