


Sample Name :GRGDSP  
 Cat. No. : RP20255  
 Lot. No. : E6047520C  
 Time Processed :8:23:15  
 Month-Day-Year Processed : 04/03/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	0	
5.00	Pumps	Pump A B.Conc	0	
20.00	Pumps	Pump A B.Conc	30	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	0	
40.00	Pumps	Pump A B.Conc	0	
40.01	Controller	Stop		

<<Column Performance>>  
 <Detector A>  
 Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment: SS-CM-0286



1 Det.A Ch1 / 220nm  
 Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	7.889	86046	5406	1.614
2	14.625	40	23	0.001
3	14.756	2256	382	0.042
4	14.883	3636	553	0.068
5	15.223	17469	2993	0.328
6	15.417	5125908	630767	96.174
7	18.510	10847	733	0.204
8	20.369	22600	678	0.424
9	20.962	12527	575	0.235
10	21.822	13344	242	0.250
11	23.411	13995	1860	0.263
12	24.811	21144	2303	0.397
Total		5329814	646515	100.000