

Sample Name : LCMV GP1 33 - 41
 Cat No. : RP20091
 Lot No. : PE0191906
 Time Processed : 21:18:26
 Month-Day-Year Processed : 06/04/19

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	15
25.00	Pumps	Solvent B Conc.	75
25.01	Pumps	Solvent B Conc.	95
31.00	Pumps	Solvent B Conc.	95
31.01	Pumps	Solvent B Conc.	15
40.00	Pumps	Solvent B Conc.	15
40.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ17010508

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	9.320	6977	1063	0.064
2	9.641	14071	1820	0.129
3	9.811	1545	315	0.014
4	10.104	103492	7565	0.945
5	10.399	42560	6508	0.389
6	10.678	10471647	1456072	95.636
7	11.057	309159	30070	2.824
Total		10949450	1503413	100.000