

Sample Name :Exendin (9-39)
 Cat No :RP10872
 Lot No :E5451521A
 Time Processed :7:40:32 AM
 Month-Day-Year Processed :01/23/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
27.00	Pumps	Pump A B.Conc	95
27.01	Pumps	Pump A B.Conc	5
35.00	Pumps	Pump A B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: GK11010011

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	10.443	1058	160	0.018
2	11.265	1194	125	0.020
3	13.080	1322	126	0.022
4	15.255	1577	193	0.026
5	17.706	1999	370	0.034
6	18.282	4874	466	0.082
7	18.903	122807	7603	2.060
8	19.425	4210	1063	0.071
9	19.571	38026	3357	0.638
10	20.047	5719578	500389	95.930
11	20.559	4409	1028	0.074
12	21.112	50759	3145	0.851
13	21.417	1526	227	0.026
14	22.830	3878	428	0.065
15	23.276	1140	176	0.019
16	23.667	1369	143	0.023

Peak#	Ret. Time	Area	Height	Area%
17	24.461	1155	160	0.019
18	24.904	1387	239	0.023
Total		5962269	519398	100.000