

Sample Name :Apamin
 Cat No :RP11002
 Lot No :E7793520L
 Time Processed :21:51:04
 Month-Day-Year Processed :01/09/2021


Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
31.00	Pumps	Pump A B.Conc	95
31.01	Pumps	Pump A B.Conc	5
40.00	Pumps	Pump A B.Conc	5
40.01	Controller	Stop	

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm



1 Detector A Channel 1 / 220nm

Peak Table

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	15.461	22451	2882	0.552
2	15.742	69372	9531	1.706
3	15.933	3898692	586591	95.903
4	16.196	32174	5851	0.791
5	18.850	21744	2843	0.535
6	19.626	20820	2919	0.512
Total		4065254	610618	100.000