

Sample Name :Gly-Arg-Gly-Asp-Ser
Cat. No. : RP10848
Lot. No. : PE7891812
Time Processed :9:55:40
Month-Day-Year Processed : 01/06/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
Total Flow:1 ml/min
Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	0	
5.00	Pumps	Pump A B.Conc	0	
20.00	Pumps	Pump A B.Conc	30	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	0	
40.00	Pumps	Pump A B.Conc	0	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Equipment: SS-CM-0286



1 Det.A Ch1 / 220nm
Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	4.888	16240	2108	0.385
2	5.817	57785	5688	1.370
3	6.726	102928	12438	2.441
4	6.968	4008483	209737	95.046
5	7.887	11113	666	0.264
6	8.533	10753	802	0.255
7	14.386	6563	867	0.156
8	14.667	143	28	0.003
9	15.867	168	33	0.004
10	16.013	3219	468	0.076
Total		4217395	232834	100.000