

Sample Name : LCMV gp33 – 41  
 Cat No : RP20257  
 Lot No : E5236421D  
 Time Processed : 10:18:43 PM  
 Month-Day-Year Processed : 04/18/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow: 1 ml/min  
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment: GK11010011

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	9.960	1962	377	0.031
2	11.444	43247	6772	0.690
3	12.066	1082	224	0.017
4	12.593	27754	1573	0.443
5	13.237	6132896	845702	97.849
6	13.587	11991	2785	0.191
7	14.158	4862	594	0.078
8	14.423	2549	316	0.041
9	15.175	3854	352	0.061
10	20.778	8356	934	0.133
11	21.477	3974	454	0.063
12	24.556	25219	1460	0.402
Total		6267746	861544	100.000