

Sample Name :Bombesin
 Cat No. : RP11157
 Lot No. : E1759421A
 Time Processed :10:24:13
 Month-Day-Year Processed :01/28/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ20010139

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	12.767	10086	1117	0.191
2	12.973	66439	8968	1.256
3	13.704	11508	1302	0.218
4	14.065	5038875	674852	95.262
5	14.245	66905	15899	1.265
6	15.860	18444	2791	0.349
7	16.071	1971	279	0.037
8	17.647	75249	10548	1.423
Total		5289477	715755	100.000