

Sample Name : MOG(35-55)
 Cat. No. : RP10245
 Lot. No. : E2860920B
 Time Processed: 23:51:03
 Year-Month-Day: 2020/2/27

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min

<<Detector A>>
 Wavelength Ch1 : 220 nm

<<LC Time Program>>


Time	Module	Command	Value	Comment
0.01	Pumps	Solvent B Conc.	5	
25.00	Pumps	Solvent B Conc.	65	
25.01	Pumps	Solvent B Conc.	95	
27.00	Pumps	Solvent B Conc.	95	
27.01	Pumps	Solvent B Conc.	5	
33.00	Pumps	Solvent B Conc.	5	
33.01	Controller	Stop		

<<Column Performance>>

<Detector A>
 Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010015

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	12.155	15891	1481	0.145
2	12.459	132140	9251	1.209
3	12.818	131633	10361	1.204
4	13.261	10442271	1452048	95.529
5	13.647	90230	9334	0.825
6	13.901	80121	7351	0.733
7	17.883	38705	3081	0.354
Total		10930991	1492907	100.000