

Sample Name :Substance P

Cat No. : RP10178

Lot No. : E6164321L

Time Processed :6:32:50 PM

Month-Day-Year Processed :12/27/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)

Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)

Total Flow:1 ml/min

Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
27.00	Pumps	Pump A B.Conc	95
27.01	Pumps	Pump A B.Conc	5
35.00	Pumps	Pump A B.Conc	5
35.01	Controller	Stop	

<<Column Performance>>


<Detector A>

Column :Inertsil ODS-SP 4.6 x 250 mm

Equipment: ZJ21010376

### <Chromatogram>

mV



### <Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	15.532	180812	19597	1.286
2	16.429	7910	3433	0.056
3	16.572	13795447	917982	98.137
4	17.063	2656	1818	0.019
5	17.589	38449	4449	0.274
6	18.206	32072	2214	0.228
Total		14057346	949494	100.000