

Sample Name : OVA Peptide (323-339)
 Cat. No. : RP10610
 Lot. No. : PE5001802

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	5	
40.00	Pumps	Pump A B.Conc	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Chromatogram



1 Det. A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	10.525	2089	311	0.039
2	10.709	9608	1374	0.178
3	10.867	21355	4024	0.395
4	11.068	5292265	814592	97.790
5	11.401	67561	7284	1.248
6	11.669	14187	1387	0.262
7	11.800	3087	655	0.057
8	14.335	1738	285	0.032
Total		5411891	829912	100.000