

Sample Name : β -Amyloid (1-42), rat
 Cat. No. : RP10013
 Lot. No. : PE8851807


Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump B Conc.	5	
25.00	Pumps	Pump B Conc.	65	
25.01	Pumps	Pump B Conc.	95	
31.00	Pumps	Pump B Conc.	95	
31.01	Pumps	Pump B Conc.	5	
37.00	Pumps	Pump B Conc.	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm



Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	8.305	3084	344	0.095
2	8.724	631	66	0.019
3	9.874	19735	1606	0.606
4	10.819	51727	2707	1.588
5	11.616	3137821	218674	96.358
6	12.233	43418	2786	1.333
Total		3256416	226181	100.000