

Sample Name :Angiotensin I, human  
 Cat. No. :RP11251 Lot.No. :PE8601711

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump B Conc.	5	
25.00	Pumps	Pump B Conc.	65	
25.01	Pumps	Pump B Conc.	95	
27.00	Pumps	Pump B Conc.	95	
27.01	Pumps	Pump B Conc.	5	
35.00	Pumps	Pump B Conc.	5	
35.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	14.270	8827	869	0.049
2	14.749	17725694	2158113	99.344
3	15.119	21856	2438	0.122
4	15.297	3983	571	0.022
5	15.679	16683	1573	0.093
6	16.076	37949	2860	0.213
7	17.870	27694	3380	0.155
Total		17842687	2169804	100.000