

Sample Name: GRGDSPC
 Cat No. : RP20283
 Lot No. : E2670221D
 Time Processed: 16:41:09
 Month-Day-Year Processed :04/11/2021
 Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm
 <<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	0
5.00	Pumps	Solvent B Conc.	0
20.00	Pumps	Solvent B Conc.	30
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
31.00	Pumps	Solvent B Conc.	95
31.01	Pumps	Solvent B Conc.	0
40.00	Pumps	Solvent B Conc.	0
40.01	Controller	Stop	

<<Column Performance>>


<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Equipment:ZJ19010014

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	13.161	66302	8261	0.666
2	13.984	9559469	861973	96.002
3	14.736	31627	4067	0.318
4	15.364	26358	4199	0.265
5	15.767	233541	36488	2.345
6	16.228	18073	2766	0.181
7	19.150	22236	2908	0.223
Total		9957606	920662	100.000