

Sample Name:N-Formyl-Met-Leu-Phe

Cat. No. :RP12956

Lot. No. : PE9901803

Pump A : 0.065% trifluoroacetic in 100% water (v/v)

Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)

Total Flow:1 ml/min

Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	15	
25.00	Pumps	Pump A B.Conc	75	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	15	
40.00	Pumps	Pump A B.Conc	15	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : AlltimaTM C18 4.6 x 250 mm

### Chromatogram



1 Det.A Ch1 / 220nm

### Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	13.524	6268	1400	0.124
2	14.901	7059	922	0.139
3	15.192	15056	2899	0.297
4	16.314	40777	5085	0.805
5	16.536	58884	10534	1.163
6	16.725	4910110	727330	96.959
7	17.117	17535	2970	0.346
8	17.664	8434	1338	0.167
Total		5064124	752478	100.000