

Sample Name :N-Boc-Phe-Leu-Phe-Leu-Phe
 Sample ID :U2818FH040-1
 Time Processed :13:45:31
 Month-Day-Year Processed :08/16/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
27.00	Pumps	Pump A B.Conc	95
27.01	Pumps	Pump A B.Conc	5
35.00	Pumps	Pump A B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ17010508

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	10.940	2654	304	0.182
2	11.394	1373	178	0.094
3	14.794	17931	2195	1.228
4	15.507	9541	646	0.654
5	16.236	1406623	195878	96.358
6	17.084	1414	243	0.097
7	20.683	2858	344	0.196
8	24.497	17390	2153	1.191
Total		1459784	201940	100.000