

Sample Name :Bombesin
 Cat No. :RP11157
 Lot No. :E0909019G
 Time Processed :8:43:21
 Month-Day-Year Processed :07/02/19

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
33.00	Pumps	Solvent B Conc.	5
33.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ17010508

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	14.123	7625	1283	0.058
2	14.702	52739	8377	0.400
3	15.097	3327	666	0.025
4	15.440	16622	1443	0.126
5	15.829	11485	1346	0.087
6	16.150	12798878	1858540	97.102
7	16.376	248676	35259	1.887
8	16.704	21822	2886	0.166
9	17.463	1550	225	0.012
10	18.003	1958	197	0.015
11	21.089	2765	295	0.021
12	23.098	1082	98	0.008
13	24.220	5774	605	0.044
14	24.413	6607	603	0.050
Total		13180909	1911823	100.000