

Sample Name :GRGDSPC
 Cat No. :RP20283
 Lot No. :E8975220J
 Time Processed :9:04:20 AM
 Month-Day-Year Processed :11/04/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
32.00	Pumps	Pump A B.Conc	95
32.01	Pumps	Pump A B.Conc	5
40.00	Pumps	Pump A B.Conc	5
95.00	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010324

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	9.786	22128	4293	0.430
2	10.327	30415	4236	0.590
3	10.720	65823	8107	1.278
4	11.316	22946	3835	0.445
5	11.732	4962409	913947	96.342
6	12.065	47127	5516	0.915
Total		5150847	939934	100.000