

Sample Name :GP (33-41)
 Cat. No. : RP20091
 Lot. No. : E9598819L
 Time Processed :22:27:39
 Month-Day-Year Processed :12/21/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
33.00	Pumps	Solvent B Conc.	5
33.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ17010508

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	5.415	15866	2297	0.223
2	12.039	3847	413	0.054
3	12.591	13895	2420	0.196
4	12.890	5019	784	0.071
5	13.501	2639	439	0.037
6	13.792	22381	2597	0.315
7	14.021	6932773	1205057	97.627
8	14.344	47722	8244	0.672
9	14.553	10098	1535	0.142
10	14.928	19370	3006	0.273
11	20.556	27710	3665	0.390
Total		7101319	1230457	100.000