

Sample Name :Cys-TAT(47-57)  
 Cat No.: RP20343  
 Lot No.: E6893820H  
 Time Processed :6:02:06  
 Month-Day-Year Processed :09/10/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
27.00	Pumps	Pump A B.Conc	95
27.01	Pumps	Pump A B.Conc	5
35.00	Pumps	Pump A B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm  
 Equipment: ZJ19010324

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	9.050	22552	4073	0.354
2	9.143	32159	5166	0.505
3	9.386	6166536	933688	96.859
4	9.746	10301	1965	0.162
5	10.181	64690	10404	1.016
6	10.606	21793	3729	0.342
7	11.596	41749	7031	0.656
8	11.931	6744	1028	0.106
Total		6366523	967086	100.000