

Sample Name :Apamin  
 Cat No :RP11002  
 Lot No :E2111021G  
 Time Processed :10:45:00 PM  
 Month-Day-Year Processed :08/26/2021


Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm



1 Detector A Channel 1 / 220nm

Peak Table

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	10.356	174308	30740	2.194
2	10.746	90249	14190	1.136
3	11.047	7554475	1038542	95.096
4	11.441	80294	9037	1.011
5	11.684	16791	2418	0.211
6	12.156	27935	5766	0.352
Total		7944053	1100694	100.000