


Mass Spectrum



Sample Information
Month-Day Processed : 09/26/20
Time Processed : 07:25:39 AM
Injection Volume : 0.3
Sample Name : Autocamtide 2
Cat No. : RP10271
Lot No. : E8009020I
Theoretical MW : 1527.77
Observed MW : 1527.6

Interface : ESI
Nebulizing Gas Flow : 1.5L/min
CDL Temp : 250
Block Temp : 200

Equipment : GK11010007
Interface Bias : +4.5kV
Drying Gas Flow : 5 L/min
T.Flow : 0.2ml/min
B.conc : 50%H2O/50%MeOH