

Sample Name :GRGDSPC  
 Cat No :RP20283  
 Lot No :E7708121D  
 Time Processed :3:56:14  
 Month-Day-Year Processed :04/19/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>


Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	

<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm  
 Equipment: GK11010011

<Chromatogram>



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	3.965	70076	17394	0.733
2	4.183	43071	10207	0.450
3	4.299	104819	17109	1.096
4	4.467	9266331	946831	96.864
5	5.611	82053	19973	0.858
Total		9566350	1011515	100.000