

Sample Name :Abz-FR-K (Dnp)-P-OH  
 Cat. No. :RP20115  
 Lot. No. :PE7181710

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
10.00	Pumps	Pump A B.Conc	35	
35.00	Pumps	Pump A B.Conc	95	
41.00	Pumps	Pump A B.Conc	95	
41.01	Pumps	Pump A B.Conc	5	
50.00	Pumps	Pump A B.Conc	5	
50.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	11.495	1259	336	0.107
2	12.162	1120491	170284	95.020
3	12.342	28562	4250	2.422
4	13.162	14472	2360	1.227
5	27.118	3441	587	0.292
6	31.779	10992	1643	0.932
Total		1179217	179460	100.000