

Sample Name :Cys-TAT(47-57)
 Cat No. :RP20343
 Lot No. :E1988221E
 Time Processed :16:20:08
 Month-Day-Year Processed :05/21/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

| Time | Module | Command | Value |
|-------|------------|---------|-------|
| 0.01 | Pumps | B.Conc | 0 |
| 25.00 | Pumps | B.Conc | 50 |
| 25.01 | Pumps | B.Conc | 95 |
| 32.00 | Pumps | B.Conc | 95 |
| 32.01 | Pumps | B.Conc | 5 |
| 40.00 | Pumps | B.Conc | 5 |
| 40.01 | Controller | Stop | |


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010324

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

| Peak# | Ret. Time | Area | Height | Area% |
|-------|-----------|----------|---------|---------|
| 1 | 7.465 | 27972 | 5971 | 0.154 |
| 2 | 8.075 | 17545805 | 2582415 | 96.778 |
| 3 | 8.305 | 230438 | 45647 | 1.271 |
| 4 | 8.402 | 298043 | 52766 | 1.644 |
| 5 | 8.905 | 27734 | 5639 | 0.153 |
| Total | | 18129991 | 2692438 | 100.000 |