

Sample Name :OVA Peptide (323-339)
 Cat. No. : RP10610
 Lot. No. : EE8082321C

Time Processed :13:26:38
 Month-Day-Year Processed :03/11/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)

Total Flow:1 ml/min

Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	

<<Column Performance>>


<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm

Equipment: ZJ20010140

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	8.535	10950063	1394460	98.313
2	8.746	97199	16461	0.873
3	9.211	26738	3728	0.240
4	9.412	5180	967	0.047
5	9.671	6272	779	0.056
6	10.153	2892	432	0.026
7	10.604	1399	191	0.013
8	11.381	1464	167	0.013
9	11.580	2464	193	0.022
10	12.147	24562	1888	0.221
11	12.378	6893	795	0.062
12	12.501	11184	791	0.100
13	13.109	1120	71	0.010
14	13.443	532	102	0.005
Total		11137963	1421027	100.000