

Sample Name :N-Boc-Phe-Leu-Phe-Leu-Phe
Cat No. :RP12950
Lot No. :E0060020I
Time Processed: 10:39:51 AM
Month-Day-Year Processed :9/11/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
Total Flow:1 ml/min
Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
27.00	Pumps	Pump A B.Conc	95
27.01	Pumps	Pump A B.Conc	5
35.00	Pumps	Pump A B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
Equipment:ZJ19010325

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	10.298	2675	361	0.115
2	16.007	3369	618	0.145
3	16.686	4388	716	0.188
4	16.949	2245166	306356	96.383
5	17.194	49720	7176	2.134
6	17.350	2771	766	0.119
7	17.731	21330	3712	0.916
Total		2329418	319705	100.000