

Sample Name :Bombesin
 Cat No. :RP11157
 Lot No. :E5017121F
 Time Processed :6:21:00 AM
 Month-Day-Year Processed :06/04/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm
 <<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
27.00	Pumps	Pump A B.Conc	95
27.01	Pumps	Pump A B.Conc	5
35.00	Pumps	Pump A B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010141

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	8.729	39707	4810	0.236
2	10.824	1434	176	0.009
3	11.280	2872	337	0.017
4	13.991	3229	435	0.019
5	14.496	3739	485	0.022
6	14.747	12330	937	0.073
7	15.198	3178	421	0.019
8	15.470	16394	2115	0.097
9	15.600	20833	2510	0.124
10	15.823	315055	41084	1.871
11	16.273	34012	3050	0.202
12	16.585	110932	13008	0.659
13	16.763	120107	23861	0.713
14	16.894	16112629	1682524	95.663

Peak#	Ret. Time	Area	Height	Area%
15	17.422	5305	872	0.031
16	17.895	1036	146	0.006
17	18.229	2697	350	0.016
18	18.953	7966	1108	0.047
19	19.124	3145	368	0.019
20	20.528	4286	290	0.025
21	23.008	21047	839	0.125
22	24.287	1175	161	0.007
Total		16843108	1779885	100.000