

Sample Name : OVA Peptide (257-264)
 Cat. No. :RP10611
 Lot. No. :PE0711801

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
27.00	Pumps	Pump A B.Conc	95	
27.01	Pumps	Pump A B.Conc	5	
35.00	Pumps	Pump A B.Conc	5	
35.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	15.012	1162	135	0.129
2	16.750	883434	147724	98.252
3	17.195	3374	609	0.375
4	17.824	4130	629	0.459
5	18.603	1478	191	0.164
6	19.245	1134	179	0.126
7	19.974	2698	174	0.300
8	20.782	1741	279	0.194
Total		899151	149919	100.000