

Sample Name :N-Boc-Phe-Leu-Phe-Leu-Phe
 Cat. No. : RP12950
 Lot. No. : E1161720K
 Time Processed :17:59:36
 Month-Dav-Year Processed :11/12/2020
 Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: SS-CM-0309

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	9.593	4479	619	0.264
2	13.599	14414	1031	0.849
3	13.675	16031	1047	0.945
4	13.952	19680	999	1.160
5	16.088	9746	941	0.574
6	16.403	1620836	287973	95.504
7	16.822	7480	717	0.441
8	18.855	2213	333	0.130
9	23.653	2260	197	0.133
Total		1697139	293857	100.000