

Sample Name : Atrial Natriuretic Peptide (ANP) (1-28), rat  
 Cat. No. : RP11195  
 Lot. No. : PE0631812  
 Time Processed :8:47:23  
 Month-Day-Year Processed :01/25/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
27.00	Pumps	Pump A B.Conc	95	
27.01	Pumps	Pump A B.Conc	5	
35.00	Pumps	Pump A B.Conc	5	
35.01	Controller	Stop		


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Equipment: GR11010440

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	14.766	1978	399	0.144
2	14.881	5993	873	0.436
3	15.100	1355619	177312	98.653
4	15.417	10539	1491	0.767
Total		1374129	180075	100.000