

Sample Name : SEB Domain (144-153)  
 Cat. No. : RP20430  
 Lot. No. : PE9811610


Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow: 1 ml/min  
 Wavelength: 220 nm

| Time  | Unit       | Command      | Value | Comment |
|-------|------------|--------------|-------|---------|
| 0.01  | Pumps      | Pump B Conc. | 5     |         |
| 25.00 | Pumps      | Pump B Conc. | 65    |         |
| 25.01 | Pumps      | Pump B Conc. | 95    |         |
| 31.00 | Pumps      | Pump B Conc. | 95    |         |
| 31.01 | Pumps      | Pump B Conc. | 5     |         |
| 40.00 | Pumps      | Pump B Conc. | 5     |         |
| 40.01 | Controller | Stop         |       |         |

<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

| Peak# | Ret. Time | Area    | Height | Area %  |
|-------|-----------|---------|--------|---------|
| 1     | 5.608     | 7266    | 882    | 0.159   |
| 2     | 10.267    | 6176    | 1370   | 0.135   |
| 3     | 10.508    | 4557833 | 415973 | 99.624  |
| 4     | 10.817    | 3750    | 1044   | 0.082   |
| Total |           | 4575025 | 419269 | 100.000 |