

Sample Name :Pep-1 (uncapped)  
 Cat No. : RP10130  
 Lot No. : E8230421C  
 Time Processed :12:13:17 PM  
 Month-Day-Year Processed :03/15/2021  
 Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
31.00	Pumps	B.Conc	95
31.01	Pumps	B.Conc	5
40.00	Pumps	B.Conc	5
40.00	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm  
 Equipment: ZJ19010140

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	10.526	15231	2968	0.450
2	10.700	46969	9461	1.388
3	10.979	18719	2428	0.553
4	11.129	9358	2907	0.277
5	11.263	3294036	614937	97.333
Total		3384312	632702	100.000