

Sample Name :Atrial Natriuretic Peptide (ANP) (1-28), rat  
 Cat. No. : RP11195  
 Lot. No. : PE7001710

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump B Conc.	5	
25.00	Pumps	Pump B Conc.	65	
25.01	Pumps	Pump B Conc.	95	
31.00	Pumps	Pump B Conc.	95	
31.01	Pumps	Pump B Conc.	5	
40.00	Pumps	Pump B Conc.	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	13.644	5964	810	0.541
2	13.850	2762	441	0.251
3	14.303	12861	1210	1.168
4	14.500	13664	1921	1.240
5	14.757	1054810	131429	95.759
6	15.445	3932	580	0.357
7	16.413	2262	364	0.205
8	17.291	5268	435	0.478
Total		1101523	137190	100.000