

Sample Name :GRGDSP  
 Cat No. :RP2025  
 Lot No. :E3523820I  
 Time Processed :15:09:52  
 Month-Day-Year Processed :10/01/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm  
 Equipment: ZJ19010141

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	5.332	5155895	508912	99.130
2	10.888	3205	471	0.062
3	11.983	22410	582	0.431
4	13.991	2405	234	0.046
5	16.211	3122	466	0.060
6	22.154	2350	128	0.045
7	22.450	1123	130	0.022
8	22.551	1193	117	0.023
9	23.161	1875	155	0.036
10	23.560	1150	93	0.022
11	23.947	1795	155	0.035
12	24.748	4613	374	0.089
Total		5201134	511818	100.000