

Sample Name:GRGDSPC
 Cat No. :RP20283
 Lot No. :E2827621E
 Time Processed :16:15:31
 Month-Day-Year Processed :05/10/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	0
5.00	Pumps	B.Conc	0
20.00	Pumps	B.Conc	30
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	0
35.00	Pumps	B.Conc	0
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010141

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	5.364	68679	5333	0.850
2	6.223	7806000	349424	96.589
3	8.653	9026	420	0.112
4	14.662	16724	2000	0.207
5	15.522	149482	24469	1.850
6	18.455	31786	4000	0.393
Total		8081698	385646	100.000