

Sample Name: Exendin (9-39)
 Cat. No. : RP10872
 Lot. No. : E8575819H
 Time Processed : 20:41:41
 Month-Day-Year Processed : 08/19/2019
 Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump B Conc.	5	
25.00	Pumps	Pump B Conc.	65	
25.01	Pumps	Pump B Conc.	95	
31.00	Pumps	Pump B Conc.	95	
31.01	Pumps	Pump B Conc.	0	
40.00	Pumps	Pump B Conc.	0	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Equipment: G09010878



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	13.314	19600	804	0.609
2	13.837	9046	1785	0.281
3	14.764	10145	1997	0.315
4	17.752	2330	430	0.072
5	18.068	9054	1618	0.281
6	18.450	3090574	508729	95.971
7	18.859	49832	9666	1.547
8	19.108	1890	600	0.059
9	19.258	9009	2065	0.280
10	19.570	18831	3661	0.585
Total		3220310	531355	100.000