


# Mass Spectrum



Sample Information  
Month-Day Processed : 09/08/20  
Time Processed : 09:19:51 PM  
Injection Volume : 0.3  
Sample Name : Neurotensin  
Cat No. : RP10439  
Lot No. : E3873201  
Theoretical MW : 1671.92  
Observed MW : 1672.6

Interface : ESI  
Nebulizing Gas Flow : 1.5L/min  
CDL Temp : 250  
Block Temp : 200

Equipment : GK11010007  
Interface Bias : +4.5kV  
Drying Gas Flow : 5 L/min  
T.Flow : 0.2ml/min  
B.conc : 50%H<sub>2</sub>O/50%MeOH