

Sample Name: HEL (46-61)
 Cat. No. : RP11736
 Lot. No. : E6044420C
 Time Processed: 4:21:14
 Month-Day-Year Processed :04/26/2020
 Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm
 <<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
35.00	Pumps	Solvent B Conc.	5
35.01	Controller	Stop	

<<Column Performance>>


<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Equipment:ZJ19010014

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	14.210	1014222	130965	97.924
2	15.323	5185	558	0.501
3	17.294	12417	1485	1.199
4	22.565	1499	193	0.145
5	24.274	2398	302	0.231
Total		1035720	133503	100.000