

Sample Name : Cys-TAT(47-57)
 Cat. No. : RP20343
 Lot. No. : E3153120H
 Time Processed : 12:43:44
 Month-Day-Year Processed : 09/05/2020
 Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
30.00	Pumps	Solvent B Conc.	95
30.01	Pumps	Solvent B Conc.	5
41.00	Pumps	Solvent B Conc.	5
41.00	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment : ZJ19010324

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	11.210	971	96	0.012
2	11.383	406	100	0.005
3	11.725	136264	38170	1.707
4	11.878	7775950	1211322	97.389
5	12.428	13194	1818	0.165
6	12.663	1760	353	0.022
7	13.175	2459	237	0.031
8	13.367	6054	367	0.076
9	13.550	2845	348	0.036
10	13.920	29040	3525	0.364
11	14.363	4045	600	0.051
12	15.080	8881	1166	0.111
13	16.505	1449	243	0.018
14	17.675	1124	105	0.014
Total		7984443	1258449	100.000