

Sample Name :Growth Hormone Releasing Factor (GHRF) (1-44), human
 Cat No.:RP10734
 Lot No.:E4758819I
 Time Processed :8:54:56
 Month-Day-Year Processed :09/19/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
33.00	Pumps	Solvent B Conc.	5
33.00	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment:ZJ19010324

<Chromatogram>

mV



<Peak Table>

Detector A 220nm

Peak#	Ret. Time	Area	Height	Area%
1	13.008	7878	672	0.185
2	13.350	1544	258	0.036
3	13.646	127642	9907	2.997
4	14.167	4051657	367493	95.122
5	14.650	70728	5151	1.661
Total		4259449	383481	100.000