

Sample Name: OVA Peptide (323-339)
 Cat No. : RP10610
 Lot No. : E4006321F
 Time Processed : 22:13:22
 Month-Day-Year Processed : 06/18/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

<<LC Time Program>>


Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010141

<Chromatogram>



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	8.076	2183910	380826	96.460
2	8.279	28996	7559	1.281
3	8.690	6392	700	0.282
4	9.106	1920	202	0.085
5	22.502	41430	2661	1.830
6	24.395	1410	120	0.062
Total		2264056	392068	100.000