

Sample Name:Pep-1 (uncapped)
 Cat No: RP10130
 Lot No: E4402621E
 Time Processed :19:10:41
 Month-Day-Year Processed :05/28/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
31.00	Pumps	B.Conc	95
31.01	Pumps	B.Conc	5
40.00	Pumps	B.Conc	5
40.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010140

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	12.473	270541	22182	1.380
2	12.775	18704828	1692101	95.388
3	13.371	353258	26588	1.801
4	13.917	220989	33340	1.127
5	14.603	59640	13772	0.304
Total		19609256	1787984	100.000