

Sample Name:MOG(35-55)  
 Cat No. :RP10245  
 Lot No. :E4062521G  
 Time Processed :12:21:19 PM  
 Month-Day-Year Processed :02/25/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
30.00	Pumps	Pump A B.Conc	95
32.00	Pumps	Pump A B.Conc	5
40.00	Pumps	Pump A B.Conc	5
45.00	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm  
 Equipment: ZJ19010324

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	16.654	213165	27726	1.812
2	17.045	11552168	1590701	98.188
Total		11765333	1618426	100.000